

Cryostation RF Coax Feedthroughs	Doc-106
----------------------------------	---------

An unused optical port can be converted to a feedthrough for coax. An SMA connector mates to the outside, while inside small coax is thermally lagged and brought into the cold space of the sample chamber.

